



Department of Botany
Panjab University, Chandigarh 160014, INDIA

Dr. Shalinder Kaur
Principal Investigator

Tel. (0172) 253 4095
E-mail: shalinder@pu.ac.in;
shalinderkaur@hotmail.com

98-105/ DST/ SK
August 5, 2016

M/s
.....
.....

Subject: *NIQ for the supply of EMF shielded Chamber (Qnt. 01)*

Sir,

Kindly quote your lowest rate for the supply of a **EMF shielded Chamber** as per the given details:

Specifications:

- EMF Shielded cabinet 1020x1020x1020 mm inner working chamber with modular galvanized panels.
- Knife edge manual single door for effective EMF shielding & for typical 500000 openings
- Honeycomb air vent to enable air flow in the enclosure
- Cut off frequency 18 GHz, Feed through shielded panel to penetrate the cabinet through coaxial connectors
- Chamber capable to bear temperature between 0 – 50 °C.
- Metallic Rack structure

The quotation (technical and financial separately) in a sealed cover and quoting lowest rate along with the applicable taxes and discounts for the universities should reach **Dr. Shalinder Kaur, Principal Investigator, Department of Botany, Panjab University Chandigarh, latest by August 19, 2016.**

Kindly mention “**Quotation for EMF Shielded Chamber**” on the cover of the Envelope.

Yours Sincerely,

Shalinder Kaur
Principal Investigator