

VeriFlex Blocks range 10°C (5°C zone-to-zone, with 3 VeriFlex™ zones)

Temperature accuracy $\pm 0.25^\circ\text{C}$ (35°–99.9°C)

Max block ramp rate* 4°C/sec

Max sample ramp rate* 3°C/sec

Temperature range 0°–100.0°C

Temperature uniformity: $<0.5^\circ\text{C}$ (30 sec after reaching 95°C)

Temperature calibration: Calibrated to standards traceable to the National Institute of Standards and Technology

Display interface: 6- 8-inch color TFT LCD

Warranty: Two Years

The rates quoted should be **FOR Botany Department, P. U., Chandigarh. SPECIAL DISCOUNT** for educational institutions/University teaching department may be mentioned. **The technical, financial bids and EMD be enclosed in separate sealed envelopes** duly marked and placed together in a duly sealed envelope giving our reference and due date of quotation. It should reach to **Dr. S. K. Upadhyay, Principal Investigator Department of Botany, Panjab University, Chandigarh-160 014 by due date.** Warranty both parts labour should be minimum of two years.

Thanking you,

Yours sincerely,

(S. K. Upadhyay)
Principal Investigator