

DEPARTMENT OF BOTANY, PANJAB UNIVERSITY, CHANDIGARH

Ref. 21-26/RJ/Bot
July 25, 2019

N.I.Q.
UNDER REGISTERED POST

Dear Sir,

Please quote your lowest rates in DUPLICATE, for the item (s) given below as per the specifications. Your quotation should reach the undersigned on **or before 9.8.2019**.

Orbital shaking Incubator Cum-BOD Incubator

Specifications:

Usable Space in MM: 800x800x800, Capacity: 12 cu.Ft, Volume (Ltr.) 275, No. of shelves: 2, Temperature Range: 4°C above ambient to 120°C (Resolution 0.1°C), Temp. controller: By Microprocessor Based PID Digital Temperature indicator cum controller, Shaking amplitude: 25 mm., RPM Display: Digital, Shaking speed: Variable speed from 10 RPM to 350 RPM, Temp. Accuracy: $\pm 1^\circ\text{C}$, Temp. Display: LED Display for Set Value (SV) and Process Value (PV), Relay: Solid State Electronic relay with protective heat sink., Air circulation: By forced convection system, Insulation By high density PUF/ Glass Wool, Operations: Nearly silent operation with ultra-low vibration, High voltage Protection: Through automatic Voltage Stabilizer (Optional), Electric supply: 220/230 V AC, 50/60/Hz, Shaking Platform Accommodate: Platform size 55x55 cm cap. 200 Lit., 24 Flasks of 500 ML/ 250 ML Capacity

Warranty both parts labour should be minimum of two years.

The rates quoted should be **FOR Botany Department, P. U., Chandigarh. SPECIAL DISCOUNT** for educational institutions/University teaching department may be mentioned. **The technical, financial bids and EMD be enclosed in separate sealed envelopes** duly marked and placed together in a duly sealed envelope giving our reference and due date of quotation. It should reach to **Dr. Rupali Jandrotia, Principal Investigator Department of Botany, Panjab University, Chandigarh-160 014** by due date. **Warranty both parts labour should be minimum of two years.**

Thanking you,

Yours sincerely,

(Rupali Jandrotia)
Principal Investigator
Department of Botany
Panjab University
Chandigarh